



EKOTEKS

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE



AB-0583-T

2400012418

12.24

TEST REPORT
DENEY RAPORU

Report Nr / Revision Nr : 2400012418
Customer name : LTS TEKSTİL SANAYİ VE TİC. A.Ş.
Applicant Address : ATATÜRK MH C GÜRSEL CD NO 6 5
KIRAÇ ESENYURT / İSTANBUL
Buyer name : -
Contact Person : -
Sample Accepted on : 23.12.2024
Re-submitted/re-confirmation date : -
Date of test : 24.12.2024 / 27.12.2024
Fiber Composition : Claimed to be :-
Sample Description : Nasal Aspirator & Replacement Tips
Order No : 92241104115 / 8684646781053 / 8692241104214 / 8684646781060
Model Number : Art.411 / Art.421
Previous Report No : -
Sampling : The results given in this report belong to the received sample by vendor.
End use : -
Care label : -
Decision Rule : -
Disclaimer Statement : -
Conformity Assessment : -

The Turkish Accreditation Agency (TURKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports. EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.

The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.

Seal

Date
27.12.2024

Customer Representative
YAGMUR OKUMUS

Laboratory Manager
Sevim A. RAZAK
27.12.2024



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T
2400012418
12.24

TEST RESULTS

TEST	METHOD	EVALUATION
DETERMINATION OF PHTHALATES	CPSC-CH-C1001-09.4:2018	P
DETERMINATION OF ORGANIC TIN COMPOUNDS	ISO/TS 16179:2012 (DOT)	P
CADMIUM	EN 1122:2001	P
DETERMINATION OF LEAD (Pb)	CPSC-CH- E1002.08.3:2012	P

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The declaration of conformity was given in accordance with the Simple Acceptance Decision Rule. (without considering the level of confidence and measurement uncertainty, evaluation of suitability or non-conformity based on whether the test result obtained is only within the specified limits) Tests marked (*) in this report are not included in the accreditation schedule.

Photo of the sample received



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.

AB-0583-T
2400012418
12.24

TEST RESULTS

Sample ID	Part No.	Part ID	Explanatory Information
A	1	PLASTIC ASPROTOR	
A	2	PLASTIC PIPE & THE PART AT THE END OF THE PIPE	
A	3	HARD ROUND PLASTIC PART	

Summary of Results

Parts & Results															
Parts	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
DETERMINATION OF PHTHALATES	P	P	P	-	-	-	-	-	-	-	-	-	-	-	-
DETERMINATION OF ORGANIC TIN COMPOUNDS	P	P	P	-	-	-	-	-	-	-	-	-	-	-	-
CADMIUM	P	P	P	-	-	-	-	-	-	-	-	-	-	-	-
DETERMINATION OF LEAD (Pb)	P	P	P	-	-	-	-	-	-	-	-	-	-	-	-

P Pass
F Fail
LS Lack of sample.
NA Not applicable
R Refer to retailer technologist

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T
2400012418
12.24

TEST RESULTS

DETERMINATION OF PHTHALATES

CPSC-CH-C1001-09.4:2018

Test sample is extracted in suitable organic solvent in extraction device. The extract is analyzed by GC-MS and result is given.

Tested Parts	Result (ppm)	Result (%)	Total Results (ppm)	Total Results (%)	Limit Value (ppm)	Limit Value (%)	Extraction Method	Evaluation
1+2	N.D	<0.006	< 60	<0.006	Total: 1000	Total : 0.1	-	PASS
3	N.D	<0.006	< 60	<0.006	Total: 1000	Total : 0.1	-	PASS

BBP/ Butylbenzylphthalate (85-68-7) DBP/ Dibutylphthalate (84-74-2)
DEHP/ Di-(2-ethylhexyl) phthalate (117-81-7) DCHP/ Di-cyclohexylphthalate (84-61-7)
DIBP/ Bis - iso -butyl phthalate (84-69-5) DINP/ Di-iso-nonylphthalate (28553-12-0)/ (68515-48-0)
DNHP/ Di-n-hexyl phthalate (84-75-3) DMP/ Dimethylphthalate (131-11-3)
DPP/ Di-pentylphthalate (n-, iso-, or mixed) (131-18-0) / (605-50-5) (776297-69-9) / (84777-06-0)
DEP/ Di-ethylphthalate (84-66-2) DMEP/ Di-(2-methoxyethyl) phthalate (117-82-8)
DIHP/ Di-C6-8-branched alkylphthalates (71888-89-6) DNOP/ Di-n-octylphthalate (117-84-0)
DIDP/ Di-isodecylphthalate (26761-40-0)/ (68515-49-1) DHNUP/Di-C7-11-branched alkylphthalates (68515-42-4)
DHP/ Di-hexylphthalates, branched and linear (68515-50-4)
DIHxP/ Di-iso-hexylphthalate (71850-09-4) DIOP/ Di-iso-octylphthalate (27554-26-3)
DPrP/ Di-n-propylphthalate (131-16-8) DNP/ Di-n-nonylphthalate (84-76-4)
1,2-Benzenedicarboxylic acid, di-C6-10 alkyl esters (68515-51-5)
1,2-Benzenedicarboxylic acid, mixed decyl and hexyl and octyl diesters (68648-93-1)

Total Uncertainty:17%

DETERMINATION OF ORGANIC TIN COMPOUNDS

ISO/TS 16179:2012 (DOT)

Sample is extracted with organic solvent then derivated and gas chromatographic test.

Tested Parts	Result (ppm)	Limit Value (ppm)	Evaluation
1+2	N.D	1000	PASS
3	N.D	1000	PASS

Di-n-octyltin-dichloride (DOT) (CAS No : 3542-36-7)

Total Uncertainty:9,3%

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

2400012418

12.24

TEST RESULTS

CADMIUM

EN 1122:2001

The Cadmium content in sample is separated by wet-decomposition method.Amount of cadmium is determined using ICP-MS.

Tested Parts	RESULT	REQUIREMENT	Evaluation
1+2	N.D	100	PASS
3	N.D	100	PASS

Reporting Limit : 2 ppm (0.0002 %)
< : Less than
ppm (part per million) : mg/kg

Controlled Metal :

Cadmium (Cd) (CAS No : 7440-43-9)

The sample have been determined by the Microwaver Method B.

The results have been determined by the device of Inductively Coupled Plasma - Mass Spectrometer (ICP- MS)

N.D : Not detected

Total Uncertainty:%3,40

DETERMINATION OF LEAD (Pb)

CPSC-CH- E1002.08.3:2012

Tested Parts	Result (ppm)	Limit Value (ppm)	Evaluation
1+2	N.D	500	PASS
3	N.D	500	PASS

Limit Value : It is given by the EU Regulation 1907/2006 (REACH), Annex XVII Entry 63.

Reporting Limit : 3 ppm
< : Less than
ppm (part per million) : mg/kg

Controlled Metal :

Lead (Pb) (CAS No : 7439-92-1)

The results have been determined by the device of Inductively Coupled Plasma - Mass Spectrometer (ICP- MS)

N.D (1): Not detected

total uncertainty:%26,04